



DECLARATION OF CONFORMITY

The Company

Name:

Advanced Bio-Technologies Inc.

Address:

ADVANCED BIO-TECHNOLOGIES, INC, 4830 WEST KENNEDY BLVD, SUITE 600,

TAMPA, FL 33609

With EU Authorised representative:

Name:

Alliance Pharma S.r.l.

Address:

Alliance Pharma S.r.l., Viale Restelli 5, 20124, Milan, Italy

DECLARES

Device Name:

KELO-COTE GEL

Other Trade Names:

N/A

Device Classification:

I (Rule 1)

Conformity assessment procedure:

Directive 93/42 EEC Annex VII

- meets the essential requirements stated in Annex I of the European Directive 93/42/EEC and its amendments Directive 2007/47 EC.
- is put on the market with the CE mark, according to the European Directive 93/42/EEC and its amendments till Directive 2007/47 EC.

Applicable harmonised standards applied to this device.

| Standard / Guideline Number | Name / Title |
|--------------------------------|---|
| BS EN ISO 13485:2012 | Medical Devices- Quality Management Systems – Requirements for Regulatory Purposes |
| BS EN ISO 14971:2012 | Medical Devices- Application of Risk Management to Medical Devices |
| BS EN ISO 10993-1: 2009 | Biological evaluation of medical devices Part 1: Evaluation and testing within a risk management process |
| BS EN ISO 10993-5:2009 | Biological Evaluation of Medical Devices-Part 5: Tests for In-vitro cytotoxicity |
| BS EN ISO 10993-10:2009 | Biological Evaluation of Medical Devices-Part 10: Tests for irritation and skin sensitisation |
| BS EN ISO 15223-1:2012 | Medical Devices- Symbols to be used with Medical Device labels, labelling and information to be supplied General Requirements |
| BS EN 1041: 2008 + A1:2013 | Information supplied by Manufacturers with Medical Devices |
| MEDDEV 2.7.1 | Guidelines on Medical Devices; Clinical Evaluation |
| MEDDEV 2.12/1 | Guidelines on Medical Devices; Guidelines on a Medical Devices Vigilance System |
| MEDDEV 2.12/2 | Guidelines on Medical Devices; Post Market Clinical Follow Up Studies |

The manufacturer undertakes to keep, and place at the Authorities disposal, the device technical file for a period of five years from the last date of manufacture of the product.

Date: 4/1/17

Advanced Bio-Technologies Inc. Legal Representative